

Determination of Antimonite (Sb(III)) and Antimonate (Sb(V)) Species using Hydride Generator-Quartz Flame Atomic Absorption Spectrophotometry (HG-QFAAS) Techniques

by Herlinawati Herlinawati

Submission date: 22-Jan-2021 11:07AM (UTC+0700)

Submission ID: 1491950359

File name: 3.pdf (354.73K)

Word count: 3355

Character count: 16627

16 Determination of Antimonite (Sb(III)) and Antimonate (Sb(V)) Species using Hydride Generator-Quartz Flame Atomic Absorption Spectrophotometry (HG-QFAAS) Techniques

Herlinawati
Doctor Program

B. Buchari & M. B. Amran

Faculty of Mathematics and Natural Sciences
Department of Chemistry
Institute Technology of Bandung
Indonesia

Abstract

20
The determination of antimonite (Sb(III)) and antimonate (Sb(V)) species using HG - QFAAS have been done. To increase the sensitivity of the measurements, a gas-liquid reactor - separator integrated into the HG-QFAAS system has been successfully prepared. In fact, this gas-liquid reactor-separator is able to increase the sensitivity of determination of Sb ions in solution. The optimum conditions of parameters measurement in the determination of Sb ion species with HG method are influenced by the type of acids, the acid concentrations, and the reductant concentrations have been determined, and can increase the measurement of analytical performance simultaneously. The result obtained showed that the antimony species provide a good measurement signal. The optimization results showed that the highest absorbance value can be obtained by using 5 M HCl and 1%NaBH₄. The analytical performance obtained is very good, which is shown by repeatability level has been 13)wn by it's coefficient of variance was 2,33% for the Sb(III) standard concentration 6 μg.L⁻¹ and 4,40 % for the Sb(V) standard concentration 150 μg.L⁻¹, linearity of the calibration curve are 2 - 10 μg.L⁻¹ for Sb(III) and 50 - 250 μg.L⁻¹ for Sb(V), detection limit of 0,26 μg.L⁻¹ for the Sb(III) and 8,42 μg.L⁻¹ for the Sb(V).

Keywords: Antimony species, hydride generation, antimonite (Sb(III)), antimonate (Sb(V)), HG-QFAAS.

1. Introduction

Antimony in nature is usually found in the form of stibnite ore (Sb₂S₃) and valentinite (Sb₂O₃). Antimony is used as a catalyst in the production of chlorofluorocarbons (CFCs) and is applied to the paint industry, batteries, rubber, drugs, bullets, pipes, and lighters. Very pure antimony is used to make semiconductor devices such as diodes and infrared detectors, while antimony compounds in stibnite form are used in the cosmetics industry. Antimony and its compounds can be released into the environment in the form of dust, volcanic eruptions, sea water, forest fires, and biogenic sources (Atlanta.G.A., 2010).

7
Antimony found in nature in several chemical forms as antimonite (Sb(III)) and antimonate (Sb(V)). 10) d Sb were combined with organic compounds and inorganic or colloid (Andreae, 1983). Antimonite (Sb(III)) and antimonate (Sb(V)) is the Sb species normally present in the environment and can be obtained on the plant 13) om soil containing Sb that causes health effects in humans (Wilson, SC., 2010). The antimony toxicity depends on the chemical form 17) is obtained, in which the element antimony is more toxic than salt and compounds of antimony in the trivalent form 25) enerally more toxic than the pentavalent form (Ida De Gregori, et al., 2007). The allowed maximum content of Sb in drinking water is allowed in China is 5 μg/L, while the WHO establish Sb level in safe drinking water is 20 μg/L (Wang X., 2012). The toxicity and Sb different species migration in the environment that affect the health of the community. To evaluate the toxicity and biogeochemistry Sb in the required 9) vironmental speciation analysis (Zheng J., 2000). Speciation of antimony and its compound include the determination of antimonite (Sb(III)) and antimonate (Sb(V)) and antimony organic compounds. The speciation has been done by gas chromatography, liquid chromatography and capillary electrophoresis method (B. Michalke, 1999).

Determination by HG-AAS is a sensitive analytical method for the elements As, Bi, Ge, Pb, Sb, Sn, and Te. Because of low interference, more selective, and more sensitive, Hydride Generation (HG) combined with AAS detection has been selected for the determination of elements in the form of hydrides (Kumar, A.R., 2005). Several methods of analysis specification of antimony compounds using various detection techniques have been done before, among others, the analysis of the speciation of Sb(III) and Sb(V) compounds by HG-ICP-OES method (Cabral, et al., 2008), detection of methylantimony compounds by QFAAS (Craig, et al., 1999), analysis of antimony speciation by HG-AFS method (Long deng, et al., 2001). HG-QFAAS detection techniques for antimonite (Sb(III)) and antimonate (Sb(V)) determination has been described in this paper. The optimization has been studied include the type and concentrations of acid, and concentrations of reductor (NaBH_4).

2. Materials and Methods

The instruments used in this study are: Hydride Generator (HG) and Quartz Flame Atomic Absorption Spectrophotometer (QFAAS) double beam-Avanta GB 6506.

All chemicals and reagents used in this study are: Potassium antimonyl tartrate trihydrate (Sigma-Aldrich), Potassium hexahydroxoantimonate (V) (Sigma-Aldrich), HCl 37% p.a (Merck), HNO_3 65% (Merck), H_2SO_4 98% (Merck), NaOH, NaBH_4 p.a (Merck), and aquadest.

Experimental Procedure

2.1 Measurement of Sb(III) and Sb(V) ions with HG-QFAAS

The determination of Sb(III) and Sb(V) ions is performed to obtain the optimum HG-QFAAS measurement conditions by varying the various parameters that affect the measurement results such as the concentration and type of acid. In this case the types of acids studied were HCl, HNO_3 , and H_2SO_4 . Also studied the concentration of HCl and NaBH_4 .

2.2 Types and Concentrations of Acid

In optimizing the type of acid and concentration, the concentration of NaBH_4 is fixed. Concentrations of acidic solutions were made to remain 5 M. The types of acids studied were HCl, HNO_3 , and H_2SO_4 . From the experiments that have been done, will get the type of acid that provides optimal response. This selected acid was used in subsequent studies by varying the concentration. The acid concentration was varied at concentrations of 1, 2, 3, 4, and 5 M.

2.3 NaBH_4 Concentration

After obtaining the optimum type and concentration of acid then varied NaBH_4 concentration with both parameters above made fixed. Variation of NaBH_4 concentration was 0.1; 0.5; 1; 1.5 and 2.0%.

2.4 Analytical Performance Determination of HG-QFAAS

Various measurements include measurement repeatability, linear range (calibration curve), and measurement detection limits using the optimum conditions obtained. Repeatability is determined by measuring the absorbance of standard solutions of Sb 50 and 100 mg/L repeatedly with the same measurement conditions under the optimum conditions as above. Repeatability is represented by % CV (coefficient of variance). Measurements to find the measured linear range of measurements were made by varying the concentrations of standard ion solutions of Sb(III) and Sb(V) ions, where the standard injection volume parameters, concentrations and types of acids, NaBH_4 concentrations as well as acid and NaBH_4 flow rates were fixed. The concentration variations of Sb(III) are 2, 4, 6, 8, and 10 $\mu\text{g/L}$ and Sb(V) used are 50, 100, 150, 200, and 250 $\mu\text{g/L}$. From this experiment we will get the equation of regression line for the measurement of Sb(III) and Sb(V) with the concentration of standard series solution. The detection limit is determined by measuring the smallest absorbance pricing that can still be determined and differentiated from the signal given by the blank with multiple measurements. The detection limit is expressed as a standard signal comparison (S) against the blank signal (N) or $S/N = 3$.

3. Results and Discussion

3.1 Determination Sb Ions with AAS Method

Sb(III) and Sb(V) metal ions can be determined by measuring directly with AAS. In this study, the calibration curves Sb(III) and Sb(V) using water solvent can be seen in Figure 1 and Figure 2.




Figure 1: Calibration curve of Sb(III) using water solvent with AAS method




Figure 2: Calibration curve of Sb(V) using water solvent with AAS method

From the results of measurement of calibration curve in Figure 1. and Figure 2. obtained linear range of concentration Sb(III) and Sb(V) are 20 - 100 mg/L, obtained regression equations are $y = 0,0054 x + 0,1024$ and $y = 0,0054 x + 0.1056$ with correlation coefficient (R^2) for Sb(III) = 0.9792 and correlation coefficient (R^2) for Sb(V) = 0.9792. The calibration curves Sb(III) and Sb(V) using HCl solvent can be seen in Figure 3. and Figure 4.




Figure 3: Calibration curve of Sb(III) using HCl solvent with AAS method




Figure 4: Calibration curve of Sb(V) using HCl solvent with AAS method

From the result of measurement of calibration curve in Figure 3. and Figure 4. obtained linear range of concentration Sb(III) and Sb(V) are 20 - 100 mg/L, obtained regression equations are : $y = 0,0055 x + 0,0071$ and $y = 0,0051 x + 0.1405$ with correlation coefficient (R^2) for Sb(III) = 0.9809 and correlation coefficient (R^2) for Sb(V) = 0.9784.

From the result of measurement of calibration curve of Sb(III) and Sb(V) ions with AAS directly obtained a small absorption when compared with measurement of Sb calibration curve with HG-QFAAS technique (in Fig. 8 and Fig. 9) the obtained absorption is bigger thus affecting the detection limit and measurement sensitivity.

3.2 Optimization Type of Acid with HG-QFAAS Techniques

Sb(III) and Sb(V) ion measurements can be determined by hydride generation (HG) method to increase the detection limit and measurement sensitivity. In this method the Sb ion is first derivatized into a hydride form and subsequently atomized in quartz cells in AAS. The HG-QFAAS techniques in this study is used for the determination of antimony species (Sb(III) and Sb(V)).

The result obtained showed that the antimony species provide a good measurement signal. Measurements were made by varying the three types of acids are HCl, H₂SO₄, and HNO₃ with the same concentration of 5 M to obtain the most optimization type of acid in the hydride formation process. The formed hydride by using HCl, H₂SO₄, and HNO₃ which produces the highest absorbance value can be obtained by using HCl. It can be seen in Figure 5.




Figure 5: Influence type of acid on the absorption Sb(III) and Sb(V) using HG-QFAAS techniques

According to Figure 5. can be seen that the absorbance of Sb(III) obtained is higher than the absorbance of Sb(V), this is due to the reaction of the formation of hydride Sb(III) to SbH₃ required only 1 reaction stage while the hydride Sb(V) to SbH₃ required 2 stages of reaction so less efficient.

3.3 Optimization of HCl Concentrations with HG-QFAAS Techniques

Based on various of HCl concentration which result the highest absorbance value can be obtained by using 5 M HCl, it can be seen in Figure 6.




Figure 6: The influence of HCl concentration on the absorption Sb(III) and Sb(V) using HG-QFAAS techniques

From Figure 6, shows that concentrations of HCl below 5 M the formation reaction hydride of Sb is not all Sb ions form SbH_3 . In this research, the most optimum absorbance values are obtained at 5 M HCl concentration. This shows that the greater the concentration of HCl, the greater of the absorbance.

3.4 Influence of NaBH_4 Concentration

Various concentrations of reductor (NaBH_4) performed to obtain optimization absorbance value Sb(III) and Sb(V) can be seen in Figure 7. To reduce Sb ions to SbH_3 , NaBH_4 is used as reductant. This reduction reaction in an acidic condition when the Sb solution has been mixed with the reductant. To determine the effect of NaBH_4 concentration, the variation of NaBH_4 concentration 0,1% - 2% using the optimum acid and concentration of 5 M HCl can be seen in Figure 7.




Figure 7: The influence of NaBH_4 concentration (%) on the absorption Sb(III) and Sb(V) using HG-QFAAS techniques

From the Figure 7, can be seen that the highest absorption Sb(III) and Sb(V) using 1% of NaBH_4 . This is supported by calibration curve for Sb(III) and Sb(V) in Figure 7. From Figure 7 can be seen that the optimized absorbance of Sb(III) and Sb(V) were 1% NaBH_4 concentration. This is due to the concentration of NaBH_4 below 1% absorbance is not maximized.

At higher concentrations of NaBH_4 does not increase significantly in absorbance values, but tends to decrease. Because of the higher concentration of NaBH_4 make the reaction is too fast, so there is a certain amount of water vapor that enters the quartz cell in the AAS can interfere its atomization process. The amount of H_2 is produced more so that dilution can occur in the gas-liquid separator. So for the next use of NaBH_4 solution with concentration of 1% as reductant.

3.5 Analytical Performance of HG-QFAAS

Based on the optimum conditions obtained on the determination of Sb(III) and Sb(V) by HG-QFAAS technique can be seen in Table 1. Furthermore, it can be determined analytical performance that is determination of repeatability, linearity, and detection limits of measurements.

Table 1: Condition of the optimum parameters of Sb ions measured by HG-QFAAS technique

Parameter measurement	Measurement results
The size of the gas-liquid separator reactor	9,0 x 3,0 cm
Type of acid	HCl
Concentration of acid	5 M
Concentration of NaBH ₄	1% (NaOH 1%, b/v)

3.5.1 Determination of repeatability

Repeatability measurements can be determined by the relative standard deviation. The smaller the standard deviation relative is indicated by the better repeatability measurements. Repeatability indicated by % CV (coefficient of variance). From the results obtained repeatability (% CV for Sb(III) = 2,33%; Sb(V) = 4,4%). % CV relatively small shows that this measurement method having high precision and can be used for micro analysis.

3.5.2 Linearity of measurement

Linearity measurements can be shown on a calibration curve Sb(III) and Sb(V) obtained from the regression linear curve that is the correlation coefficient (R^2) which is close to the value 1. Figure 6 shows the area of the linear range of concentration Sb(III) are 2 -10 $\mu\text{g.L}^{-1}$ obtained regression equation : $y = 0.0625x + 0.073$ with a correlation coefficient (R^2) = 0.9936. While the linear range of concentration Sb(V) are 50-250 $\mu\text{g.L}^{-1}$ obtained regression equation : $y = 0.0034x + 0.0027$ with correlation coefficients (R^2) = 0.9942. The measurement calibration curve Sb(III) and Sb(V) can be seen in Figure 8 and Figure 9.




Figure 8: The calibration curve of determination of Sb(III) by HG-QFAAS techniques




Figure 9: The calibration curve of determination of Sb(V) by HG-QFAAS techniques

From figure 8 and figure 9 it can be seen the linear regression for Sb(III) and Sb(V) obtained are $y = 0.0625x + 0.073$ and $y = 0.0034x + 0.0027$ which shows the linearity of the measurements.

3.5.3 Detection Limit of Measurement

Detection limits are determined by the smallest absorbance measurements that can still be determined and distinguished from the signals provided by the blank with multiple measurements. This detection limit shows the smallest concentration of analyte that can still be detected. From the experimental results the detection limit for the determination of Sb(III) with the HG-QFAAS technique is $0.26 \mu\text{g.L}^{-1}$. While the detection limit for determination of Sb(V) by HG-QFAAS techniques is $8.42 \mu\text{g.L}^{-1}$. The detection limit indicates the smallest amount of analyte concentration that can be detected. From the results obtained the detection limit for the determination of Sb(III) and Sb(V) with HG-QFAAS techniques were $0.26 \mu\text{g.L}^{-1}$ and $8.42 \mu\text{g.L}^{-1}$ respectively.

4. Conclusion

The optimization results for the determination of Sb(III) and Sb(V) with HG-QFAAS techniques showed that the highest absorbance value can be obtained by using 5 M HCl and 1% of NaBH_4 . A good analytical performance has been obtained as indicated by repeatability (% CV for Sb(III) = 2.33%; Sb(V) = 4.4%) the linearity of the calibration curve ($2-10 \mu\text{g.L}^{-1}$ for Sb(III) and $50-250 \mu\text{g.L}^{-1}$ for Sb(V)) and limit of detection for Sb(III) and Sb(V) are $0.26 \mu\text{g.L}^{-1}$ and $8.42 \mu\text{g.L}^{-1}$ respectively.

References

- Atlanta, GA. (2010). Toxicological Profile for Antimony, US agency for toxic substances and diseases registry : Available online : <http://www.atsdr.cdc.gov/toxprofiles/tp23.html>.
- B. Michalke & P. Schramel. (1999). Antimony Speciation in Environmental Samples by Interfacing Capillary Electrophoresis on-line to an Inductively Coupled Plasma Mass Spectrometer, Vol. 834, no.1-2, Journal of Chromatography A, pp.341-348.
- Cabral, L.M., Vania, N.M.J., & Luiza, R.S.D. (2008). Speciation of Antimony (III) and Antimony (V) using Hydride Generation for Meglumine Antimoniate Pharmaceutical Formulations Quality Control, Mem. Inst. Oswaldo Cruz vol.103, no. 2, Rio de Janeiro, on-line version ISSN 1678-8060.
- Craig, P.J., S.N. Forster., R.O Jenkins & D. Miller. (1999). An Analytical Method for the Detection of MethylAntimony Species in Environmental Matrices :MethylAntimony Levels in Some UK Plant Material, Journal of The Analyst, 124, 1243-1248.
- Deng, T L., Yu-Wei C., & Nelson B. (2001). Antimony Speciation at Ultra Trace Levels using Hydride Generation Atomic Fluorescence Spectrometry and 8-hydroxyquinoline as an Efficient Masking Agent, Elsevier, Analytica Chimica Acta, 432, 293-302.
- He M, Wang X, Wu F, & Fu Z. (2012). Antimony Pollution in China, Science Total Environment, 421-422, 41-50.
- Ida De Gregori. (2007). Speciation Analysis of Antimony in Marine Biota by HPLC-(UV)-HG-AFS :Extraction Procedures and Stability of Antimony Species, Elsevier, Science Direct, 73, Talanta, 458-465.
- Kumar, A.R. & Riyazuddin P. (2005). Mechanism of Volatile Hydride Formation and Their Atomization in Hydride Generation Atomic Absorption Spectrometry, Analytical Sciences, 21, 1401 -1410.
- M.O. Andreae, in : C.S. Wong, E. Boyle, K.W. Bruland, J.D. Burton, & E.D. Goldberg (Eds). (1983). Trace Metals in Sea Waters, Plenum Press, New York, p.1.
- Wilson SC, Lockwood PV, Ashley PM, & Tighe M. (2010). The Chemistry and Behaviour of Antimony in the Soil Environment with Comparisons to Arsenic : A Critical Review, 158, Environment Pollute, 1169-81.
- Zheng J, Masaki O & Naoki F. (2000). Antimony Speciation in Environmental Samples by Using HPLC Coupled to ICP-MS, Vol.16, Journal of Analytical Sciences, The Japan Society for Analytical Chemistry.

Determination of Antimonite (Sb(III)) and Antimonate (Sb(V)) Species using Hydride Generator-Quartz Flame Atomic Absorption Spectrophotometry (HG-QFAAS) Techniques

ORIGINALITY REPORT

17%

SIMILARITY INDEX

%

INTERNET SOURCES

17%

PUBLICATIONS

%

STUDENT PAPERS

PRIMARY SOURCES

- 1** Neri, Thaís S., Dayene C. Carvalho, Vanessa N. Alves, and Nívia M. M. Coelho. "Noteworthy Method for Direct Determination of SbIII and Total Inorganic Antimony in Natural Waters", *Journal of the Brazilian Chemical Society*, 2015.
Publication 3%
- 2** Nalan Ozdemir, Mustafa Soylak, Latif Elci, Mehmet Dogan. "Speciation analysis of inorganic Sb(III) and Sb(V) ions by using mini column filled with Amberlite XAD-8 resin", *Analytica Chimica Acta*, 2004
Publication 2%
- 3** Lúcio Mendes Cabral, Vânia Neves Moreira Juliano, Luiza Rosario Sousa Dias, Camila Braga Dornelas et al. "Speciation of antimony (III) and antimony (V) using hydride generation for meglumine antimoniate pharmaceutical formulations quality control", *Memórias do Instituto Oswaldo Cruz*, 2008 1%

4

Tayyebeh Madrakian, Elaheh Bozorgzadeh. "Spectrophotometric determination of Sb(III) and Sb(V) in biological samples after micelle-mediated extraction", Journal of Hazardous Materials, 2009

Publication

1%

5

Smichowski, Patricia. "Fractionation and Speciation Analysis of Antimony in Atmospheric Aerosols and Related Matrices", Speciation Studies in Soil Sediment and Environmental Samples, 2013.

Publication

1%

6

Cabon, J.Y.. "Determination of major antimony species in seawater by continuous flow injection hydride generation atomic absorption spectrometry", Analytica Chimica Acta, 20040223

Publication

1%

7

Zhefeng Fan. "Determination of antimony(III) and total antimony by single-drop microextraction combined with electrothermal atomic absorption spectrometry", Analytica Chimica Acta, 2007

Publication

1%

8

Xiao-Dan Wu, Jin-Ming Song, Xue-Gang Li, Hua-Mao Yuan, Ning Li. "Behaviors of dissolved

1%

antimony in the Yangtze River Estuary and its adjacent waters", Journal of Environmental Monitoring, 2011

Publication

9

Rajmund Michalski, Sebastian Szopa, Magdalena Jabłońska, Aleksandra Łyko. "Application of Hyphenated Techniques in Speciation Analysis of Arsenic, Antimony, and Thallium", The Scientific World Journal, 2012

Publication

10

Zhaofeng Ge, Chaoyang Wei. "Simultaneous Analysis of SbIII, SbV and TMSb by High Performance Liquid Chromatography–Inductively Coupled Plasma–Mass Spectrometry Detection: Application to Antimony Speciation in Soil Samples", Journal of Chromatographic Science, 2013

Publication

11

A. RAMESH KUMAR, P. RIYAZUDDIN. "Mechanism of Volatile Hydride Formation and Their Atomization in Hydride Generation Atomic Absorption Spectrometry", Analytical Sciences, 2005

Publication

12

Edwar Fuentes, Hugo Pinochet, Ida De Gregori, Martine Potin-Gautier. "Redox speciation analysis of antimony in soil extracts by hydride

<1%

1%

1%

1%

generation atomic fluorescence spectrometry",
Spectrochimica Acta Part B: Atomic
Spectroscopy, 2003

Publication

13

Vanessa Cristina Gonçalves dos Santos, Marco Tadeu Grassi, Gilberto Abate. "Speciation of Antimony(III) and Antimony(V) in Bottled Water by Hydride Generation-Inductively Coupled Plasma Optical Emission Spectrometry", Analytical Letters, 2015

<1%

Publication

14

Y Petit de PeÃ±a. "On-line determination of antimony(III) and antimony(V) in liver tissue and whole blood by flow injection â hydride generation â atomic absorption spectrometry", Talanta, 2001

<1%

Publication

15

Juliana Maria Oliveira Souza, César Ricardo Teixeira Tarley. "Preconcentration and Speciation of Sb(III) and Sb(V) in Water Samples and Blood Serum after Cloud Point Extraction Using Chemometric Tools for Optimization", Analytical Letters, 2008

<1%

Publication

16

Yong-Lai Feng, Hisatake Narasaki, Hong-Yuan Chen, Li-Ching Tian. "Speciation of antimony(III) and antimony(V) using hydride generation

<1%

inductively coupled plasma atomic emission spectrometry combined with the rate of pre-reduction of antimony", *Analytica Chimica Acta*, 1999

Publication

17

Naser Jamali Hajiani, Seyed Majid Ghaderian, Naser Karimi, Henk Schat. "A comparison of antimony accumulation and tolerance among *Achillea wilhelmsii*, *Silene vulgaris* and *Thlaspi arvense*", *Plant and Soil*, 2016

<1%

Publication

18

PHEs Environment and Human Health, 2014.

<1%

Publication

19

Sundar, Shyam, and Jaya Chakravarty. "Antimony Toxicity", *International Journal of Environmental Research and Public Health*, 2010.

<1%

Publication

20

Jian-Ping Zou, Hui-Long Liu, Jinming Luo, Qiu-Ju Xing, Hong-Mei Du, Xun-Heng Jiang, Xu-Biao Luo, Sheng-Lian Luo, Steven L. Suib. " Three-Dimensional Reduced Graphene Oxide Coupled with Mn O for Highly Efficient Removal of Sb(III) and Sb(V) from Water ", *ACS Applied Materials & Interfaces*, 2016

<1%

Publication

21

Xiaojing Long, Xin Wang, Xuejun Guo,

Mengchang He. "A review of removal technology for antimony in aqueous solution", *Journal of Environmental Sciences*, 2020

Publication

<1%

22

Louay AlSioufi, Ana M. Sánchez de la Campa, Daniel Sánchez-Rodas. "Microwave extraction as an alternative to ultrasound probe for antimony speciation in airborne particulate matter", *Microchemical Journal*, 2016

Publication

<1%

23

Wei, Chaoyang, Zhaofeng Ge, Wangsheng Chu, and Renwei Feng. "Speciation of antimony and arsenic in the soils and plants in an old antimony mine", *Environmental and Experimental Botany*, 2015.

Publication

<1%

24

Xi, Jianhong, and Mengchang He. "Removal of Sb(III) and Sb(V) from aqueous media by goethite", *Water Quality Research Journal of Canada*, 2013.

Publication

<1%

25

C.A. Chapa-Martínez, L. Hinojosa-Reyes, A. Hernández-Ramírez, E. Ruiz-Ruiz, L. Maya-Treviño, J.L. Guzmán-Mar. "An evaluation of the migration of antimony from polyethylene terephthalate (PET) plastic used for bottled drinking water", *Science of The Total*

<1%

Environment, 2016

Publication

26

Ramazan Gürkan, Merve Eser. "Application of ultrasonic-assisted cloud point extraction/flame atomic absorption spectrometry (UA-CPE/FAAS) for preconcentration and determination of low levels of antimony in some beverage samples", Journal of the Iranian Chemical Society, 2016

Publication

<1%

27

Enisa Omanovic, Kurt Kalcher. "A new chemiluminescence sensor for hydrogen peroxide determination", International Journal of Environmental Analytical Chemistry, 2005

Publication

<1%

28

Portugal, Lindomar A., Laura Ferrer, Antonio M. Serra, Douglas Gonçalves da Silva, Sérgio Luis C. Ferreira, and Víctor Cerdà. "A non-chromatographic automated system for antimony speciation in natural water exploiting multisyringe flow injection analysis coupled with online hydride generation – atomic fluorescence spectrometry", Journal of Analytical Atomic Spectrometry, 2015.

Publication

<1%

29

Francisco Pena-Pereira. "Headspace single-drop microextraction with in situ stibine

<1%

generation for the determination of antimony (III) and total antimony by electrothermal-atomic absorption spectrometry", *Microchimica Acta*, 01/2009

Publication

30

Xingwen Wang, Xuankun Li, Xu Zhang, Shahua Qian. " Speciation analysis of antimony in water samples via combined nano-sized TiO colloid Preconcentration and AFS analysis ", *J. Anal. At. Spectrom.*, 2014

Publication

<1%

Exclude quotes On

Exclude matches Off

Exclude bibliography On